

# Verification of Compliance

The device has been tested and found to comply with the stated standard, which is required by the Directive 93/42/EEC. The test results are indicated in the test report and are applicable only to the tested sample identified in the report.

## Equipment:


Type of Product : Tobii Dynavox I-12+  
Trade Mark : tobii dynavox  
Models Number : 12003611 I-12+ ETR-01, 120036XX I-12+ ETR-YY  
(X=0~9,Y=0~9,a~z,A~Z, or blank or slash; for marketing purpose only  
and no impact safety related constructions and critical components)

## Produced by:

Applicant : Tobii Technology AB  
Karlsrovägen 2D,182 53 Danderyd, SWEDEN  
Manufacturer's Name : ONYX Healthcare Inc.  
2F., No.135, Ln. 235, Baoqiao Rd., Xindian Dist., New Taipei City 231, Taiwan  
(R.O.C.)

## Applied Standards:

EN 60601-1-2 : 2007 +AC:2010  
EN 55011 : 2009 +A1:2010  
EN 61000-3-2 : 2006 +A1: 2009 +A2: 2009  
EN 61000-3-3 : 2013



.....  
*Roy Chen* Measurement Center Manager

Date: 2015/04/30



*A Test Lab Techno Corp.*

No. 140 -1, Changan Street, Bade City, Taoyuan County, Taiwan R.O.C.  
Tel : 886-3-271-0188 Fax : 886-3-271-0190

